



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

LABCAT, S. de R.L. M.I.
Galeana No. 619, Col. Moderna
Matamoros, Tamaulipas, México. C.P. 87300

*(Hereinafter called the Organization) and hereby declares that Organization is accredited
in accordance with the recognized International Standard:*

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the
operation of a laboratory quality management system
(as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

***Dimensional, Thermodynamic, Chemical, Mass, Force and Weighing Devices,
Time and Frequency, Mechanical and Electrical Calibration***
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President

Initial Accreditation Date:

May 12, 2011

Issue Date:

November 30, 2023

Expiration Date:

December 31, 2025

Accreditation No.:

70451

Certificate No.:

L23-921

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

*The validity of this certificate is maintained through ongoing assessments based on a
continuous accreditation cycle. The validity of this certificate should be
confirmed through the PJLA website: www.pjllabs.com*



Certificate of Accreditation: Supplement

LABCAT, S. de R.L. M.I.

Galeana, No. 619, Col. Moderna

Matamoros, Tamaulipas, México. C.P. 87300

Contact Name: J. Alexandro Santos Vargas Phone: (868) 149-0812

Accreditation is granted to the facility to perform the following calibrations:

Dimensional

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Calipers ^F	0.05 in to 4 in	$(350.86 + 6.94L) \mu\text{in}$	Block Gauge Mod: 516-421-12	NMX-CH-002-IMNC
	5 in to 20 in	$(360.17 + 3.99L) \mu\text{in}$	Bar Set Mod: 167-145 al 160	
Micrometer Outside ^F	0.05 in to 1 in	$(38.05 + 2.95L) \mu\text{in}$	Block Gauge Mod: 516-421-12	NMX-CH-99-SCFI
	5 in to 20 in	$(29.74 + 3.86L) \mu\text{in}$	Bar Set Mod: 167-145 to 160	
Depth Micrometer ^F	1 in to 8 in	$(220 + 35L) \mu\text{in}$	Comparison Block Gauge	NMX-CH-149-SCFI
Indicators ^F	0.5 in to 2 in	$(90 + 6L) \mu\text{in}$	Mitutoyo Mod: 516-421-12	
	0.000 1 in to 1 in	$(80 + 6L) \mu\text{in}$	Head Micrometer Mitutoyo Mod: 0-1"	
Set Pin Gage ^{FO}	0.011 to 0.06	$(52.92 + 73.47L) \mu\text{in}$	Micrometer "1"	Euramet cg-6
	0.061 to 0.25	$(54.21 + 51.06L) \mu\text{in}$		
	0.501 to 0.625	$(32.23 + 42.91L) \mu\text{in}$		
Height Gauge ^F	0.6 in to 24 in	$(635.72 + 5.31L) \mu\text{in}$	Block Gauge, Bar Set Mitutoyo Mod: 516-421-12 Mod: 167-145 al 160	JIS B 7517
Rules ^F	2.54 mm to 1 016 mm	0.8 mm	Graduate Rule Mitutoyo Mod: 182-309	JID B 7516
Glass Scale ^F	2.54 mm to 300 mm	0.8 mm		
Optical Comparator ^F X Axis Linearity Y Axis Linearity	12 in maximum	520 μin	Glass Rule Mod: 172-162 Comparison Block Gauge Mitutoyo Mod: 516-421-12	JIS B 7184
	12 in maximum	520 μin		
Optical Comparator ^F Squareness of Y axis to X axis	8 in of Y travel or Maximum Y axis travel if Maximum is less than 8 in	0.005 7°	Master Square Comparison Block Gauge Mitutoyo Mod: 516-421-12	
Optical Comparator Angularity ^F	Up to 180°	0.15°	Angular Reticle Mitutoyo Comparison Block Gauge Mitutoyo Mod: 516-421-12	



Certificate of Accreditation: Supplement

LABCAT, S. de R.L. M.I.

Galeana, No. 619, Col. Moderna

Matamoros, Tamaulipas, México. C.P. 87300

Contact Name: J. Alexandro Santos Vargas Phone: (868) 149-0812

Accreditation is granted to the facility to perform the following calibrations:

Dimensional

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Optical Comparator Magnification ^O	10X	0.03 %	Comparison Block Gauge Mitutoyo Mod: 516-421-12	NMX-CH-3650
	20X	0.03 %		
	31.5X	0.03 %		
	62.5X	0.03 %		
	50X	0.04 %		
Angle Block and Protactor ^F	Up to 90°	0.2°	Westward D Mini Mag Protactor Electronic	CEM DI-003
Thickness Gauge ^F	23 μ m to 1 457 μ m	2 μ m	Plastic Shims Defelsko Mod: Set	ASTM D7091
Granite Surface Plate Repeat Only ^F	0.002 μ m	80 μ m	Repeat Reading Gauge Starrett with Electronic Indicator	GGG-P-463c
Thread Plug Gage (Pitch Diameter) ^F	3-48 to 1-14	(53.96 + 3.21L) μ m	Micrometer 1'' Three Wires Gages	ASME B1.1 ASME B1.13M
Plug Gage Go, No Go ^F	0.085 5 in to 0.960 9 in	(53.96 + 3.21L) μ m		
Thread Ring Gages (Pitch Diameter) ^F	10-24 to 4-4 (Pitch Diameter 0.161 9 in to 3.834 2 in)	(2.4 + 6.26 x 10 ⁻¹ L) μ m	MAHR 828 Universal Length Machine MAHR Internal Thread Probes (Accessories)	ASME B1.1, ASME B1.13M
Gages Block, Steel Grade 0,1,2 ^F	0.05 in to 4 in	(2.35 + 1.09L) μ m	Gauge Block Set Grade 00 MAHR 828 Universal Length Machine	ASME B89.1.9
Cylindrical Ring Gages ^F	0.2 in to 20 in	(2.17 + 1.68L) μ m	MAHR 828 Universal Length Machine, Internal Probe	ASME B89.1.5
Standard and Measuring Rods to Micrometer Setting ^{FO}	1 in to 17 in	(64.49 + 4.13x 10 ⁻¹ L) μ m	Gauge Block Set Grade 00 Gauge Block Set Grade 0 Digital Indicator	MAHR MILLITAST 1085 BS 5317

Thermodynamic

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Hygrometers ^F	10 % RH to 90 % RH	2 % of reading	Temperature and Humidity Recorder and Humidity Chamber Generator	CEM TH-007
Humidity Tester ^F	10 % RH to 90 % RH	2 % of reading		
Humidity Chamber ^F	10 % RH to 90 % RH	2 % of reading		



Certificate of Accreditation: Supplement

LABCAT, S. de R.L. M.I.

Galeana, No. 619, Col. Moderna

Matamoros, Tamaulipas, México. C.P. 87300

Contact Name: J. Alexandro Santos Vargas Phone: (868) 149-0812

Accreditation is granted to the facility to perform the following calibrations:

Thermodynamic

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Oven ^{FO}	30 °C to 150 °C	0.089 % of reading	Omega CL3515R Temperature Simulator and Reader, Type K Thermal Probe Specification and Temperature- Electromotive Force	CENAM Technical Guide
	151 °C to 530 °C	0.026 % of reading		
	531 °C to 810 °C	0.026 % of reading		
	811 °C to 1 345 °C	0.026 % of reading		
Infrared Pyrometers ^{FO}	50 °C to 500 °C	0.17 °C	Black Body Reed BX-500	

Mass, Force and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Force -Tension ^O	0.001 N to 49 N	0.2 % of reading	Dead Weight F Rice Lake Mod: 1 MG-5KG	ISO 7500-1
	2 kN to 22 kN	0.5 % of reading	Force Gauge Dillon Mod: EDxtreme	
	4 kN to 44 kN	0.3 % of reading	Load Cell Transducer Techniques Mod: SW0-20K	
Force - Compression ^O	0.001 N to 49 N	0.2 % of reading	Dead Weight F Rice Lake Mod: 1MG-5KG	
	2 kN to 22 kN	0.5 % of reading	Force Gauge Dillon Mod: EDxtreme	
	4 kN to 44 kN	0.3 % of reading	Load Cell Transducer Techniques Mod: SW0-20K	
Scales and Balances ^O	1 mg to 500 mg (Res.= 0.1 mg)	$(2.57 \times 10^{-1} + 1.35 \times 10^{-3} \text{Wt}) \text{ mg}$	Class M1 Weighs Rice Like Mod:	Euramet_cg-18
	500 mg to 500 g (Res.= 1 mg)	$(1.5 \times 10^{-3} + 5.49 \times 10^{-5} \text{Wt}) \text{ g}$		
	500 g to 5 kg (Res.= 10 g)	12 g		
	5 kg to 450 kg (Res.= 0.5 kg)	580 g	Class M1 Weights Esher	



Certificate of Accreditation: Supplement

LABCAT, S. de R.L. M.I.

Galeana, No. 619, Col. Moderna

Matamoros, Tamaulipas, México. C.P. 87300

Contact Name: J. Alexandro Santos Vargas Phone: (868) 149-0812

Accreditation is granted to the facility to perform the following calibrations:

Time and Frequency

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Clock ^{FO}	9 h: 59 min: 599 s	0.6 s	Digital Stopwatch	NIST 960-12
Chronometer ^{FO}	23 h: 59 min: 599 s	5.3 s		
Timers ^{FO}	23 h: 59 min: 599 s	5.3 s		
Tachometers ^{FO}	10 rpm to 14 000 rpm	0.05 % of reading	Speed Meters and rpm Laser Photo/Contact	ASTM 432-B
Equipment to Output Frequency ^{FO}	0.1 Hz to 1 kHz	0.009 % of reading	Frequency Counter BK Precision 1856D	Euramet cg-07 Euramet cg-15
	1.1 kHz to 10 kHz	0.001 5 % of reading		
	10.1 kHz to 100 kHz	0.000 15 % of reading		
	100.1 kHz to 1 MHz	0.015 % of reading		
	1.1 MHz to 10 MHz	0.001 5 % of reading		
	10.1 MHz to 100 MHz	0.001 5 % of reading		
	100.1 MHz to 3.5 GHz	0.004 % of reading		

Mechanical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Indirect Verification of Rockwell Hardness Tester HRC ^O	20 HRC to 39 HRC	0.28 HRC	Rockwell Standardized Hardness Test Block	ASTM E18
	40 HRC to 59 HRC	0.52 HRC		
	60 HRC to 68 HRC	0.21 HRC		
Indirect Verification of Rockwell Hardness Tester HRBW ^O	10 HRBW to 50 HRBW	0.38 HRBW		
	51 HRBW to 79 HRBW	0.49 HRBW		
	80 HRBW to 100 HRBW	0.42 HRBW		
Pressure Gauge ^F	100 psi to 1 000 psi	0.25 % of reading	Omega, Pressure Gauge Mod: DPG8001-1K	ASME B40.100
	1 000 psi to 10 000 psi	0.4 % of reading	Cristal, Pressure Gauge Mod: M1-10KPSI	
Vacuum Gage ^{FO}	1 inHg to 25 inHg	0.25 % of reading	Cristal, Pressure Gauge Mod: M1-10KPSI	
Torque Wrench ^F	60 lbf·ft to 600 lbf·ft	1.2 % of reading	Torque Tester Sturtevant Richmond Mod: 600	ISO 6789-2



Certificate of Accreditation: Supplement

LABCAT, S. de R.L. M.I.

Galeana, No. 619, Col. Moderna

Matamoros, Tamaulipas, México. C.P. 87300

Contact Name: J. Alejandro Santos Vargas Phone: (868) 149-0812

Accreditation is granted to the facility to perform the following calibrations:

Mechanical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Kinematic Viscosity Zahn Cups No. 2 ^F	19 cSt a 156 cSt	0.15 % of reading	Canon Certified Viscosity Standard C60 and C200 Chronometer Casio Thermohygrometer Fluke 971	ASTM D4212
Kinematic Viscosity Zahn Cups No. 3 ^F	64 cSt a 596 cSt	0.47 % of reading		
Kinematic Viscosity Zahn Cups No. 4 ^F	79 cSt a 784 cSt	0.47 % of reading		
Kinematic Viscosity Zahn Cups No. 5 ^F	161 cSt a 1 401 cSt	0.47 % of reading		

Chemical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
pH Meter ^F	4 pH to 10 pH	0.035 pH	pH Calibrations Buffers	NMX-CH-068
Conductivity Fixed Points ^F	84 mS/cm 1 413 mS/cm	0.5 mS/cm 0.5 mS/cm	Buffer, Analytic Solution	

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Electrical Simulation of Temperature Calibration, Indication, and Control Equipment used with RTD Pt 385, 100 Ω ^{FO}	-200 °C to -80 °C	0.026 % of reading	Fluke 5522A Electrical Simulation of RTD Output	ASTM E220
	-80 °C to 0 °C	0.079 % of reading		
	0 °C to 100 °C	0.11 % of reading		
	100 °C to 300 °C	0.03 % of reading		
	300 °C to 400 °C	0.022 % of reading		
	400 °C to 630 °C	0.019 % of reading		
	630 °C to 800 °C	0.026 % of reading		
Electrical Simulation of Temperature Calibration, Indication, and Control Equipment used with RTD Pt 3926, 100 Ω ^{FO}	200 °C to -80 °C	0.05 °C		
	-80 °C to 0 °C	0.06 °C		
	0 °C to 100 °C	0.08 °C		
	100 °C to 300 °C	0.09 °C		
	300 °C to 400 °C	0.1 °C		
	400 °C to 630 °C	0.12 °C		



Certificate of Accreditation: Supplement

LABCAT, S. de R.L. M.I.

Galeana, No. 619, Col. Moderna

Matamoros, Tamaulipas, México. C.P. 87300

Contact Name: J. Alejandro Santos Vargas Phone: (868) 149-0812

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Electrical Simulation of Temperature Calibration, Indication, and Control Equipment used with RTD Pt 3916, 100 Ω^{FO}	-200 °C to -190 °C	0.29 °C	Fluke 5522A Electrical Simulation of RTD Output	ASTM E220
	-190 °C to -80 °C	0.05 °C		
	-80 °C to 0 °C	0.06 °C		
	0 °C to 100 °C	0.07 °C		
	100 °C to 260 °C	0.07 °C		
	260 °C to 300 °C	0.08 °C		
	300 °C to 400 °C	0.09 °C		
	400 °C to 600 °C	0.09 °C		
	600 °C to 630 °C	0.24 °C		
	-80 °C to 0 °C	0.03 °C		
	0 °C to 100 °C	0.05 °C		
	100 °C to 260 °C	0.05 °C		
	260 °C to 300 °C	0.13 °C		
	300 °C to 400 °C	0.14 °C		
	400 °C to 600 °C	0.14 °C		
	600 °C to 630 °C	0.16 °C		
Electrical Simulation of Temperature Calibration, Indication, and Control Equipment used with RTD Pt 385, 500 Ω^{FO}	0 °C to 100 °C	0.06 °C	Fluke 5522A Electrical Simulation of RTD Output	ASTM E220
	100 °C to 260 °C	0.07 °C		
	260 °C to 300 °C	0.08 °C		
	300 °C to 400 °C	0.08 °C		
	400 °C to 600 °C	0.09 °C		
	600 °C to 630 °C	0.1 °C		
Electrical Simulation of Temperature Calibration, Indication, and Control Equipment used with RTD Pt 385, 1 000 Ω^{FO}	-200 °C to -80 °C	0.016 % of reading		
	-80 °C to 0 °C	0.049 % of reading		
	0 °C to 100 °C	0.064 % of reading		
	100 °C to 260 °C	0.014 % of reading		
	260 °C to 300 °C	0.017 % of reading		
	300 °C to 400 °C	0.016 % of reading		
	400 °C to 600 °C	0.011 % of reading		
	600 °C to 630 °C	0.05 % of reading		



Certificate of Accreditation: Supplement

LABCAT, S. de R.L. M.I.

Galeana, No. 619, Col. Moderna

Matamoros, Tamaulipas, México. C.P. 87300

Contact Name: J. Alexandro Santos Vargas Phone: (868) 149-0812

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Electrical Simulation of Temperature Calibration, Indication, and Control Equipment used with RTD Pt Ni 385, 120 Ω^{FO}	-80 °C to 0 °C	0.07 °C	Fluke 5522A Electrical Simulation of RTD Output	ASTM E220
	0 °C to 100 °C	0.08 °C		
	100 °C to 260	0.15 °C		
Electrical Simulation of Temperature Calibration, Indication, and Control Equipment used with RTD Cu 427, 10 Ω^{FO}	-100 °C to 260 °C	0.35 °C		
Electrical Simulation of Temperature Calibration, Indication, and Control Equipment used with Thermocouple Type B ^{FO}	600 °C to 800 °C	0.49 °C	Fluke 5522A Electrical Simulation of Thermocouple Output	
	800 °C to 1 000 °C	0.39 °C		
	1 000 °C to 1 550 °C	0.36 °C		
	1 550 °C to 1 820 °C	0.31 °C		
Electrical Simulation of Temperature Calibration, Indication, and Control Equipment used with Thermocouple Type C ^{FO}	0 °C to 150 °C	0.27 °C		
	150 °C to 650 °C	0.23 °C		
Electrical Simulation of Temperature Calibration, Indication, and Control Equipment used with Thermocouple Type C ^{FO}	650 °C to 1 000 °C	0.27 °C		
	1 000 °C to 1 800 °C	0.44 °C		
	1 800 °C to 2 316 °C	0.73 °C		
Electrical Simulation of Temperature Calibration, Indication, and Control Equipment used with Thermocouple Type E ^{FO}	-250 °C to -100 °C	0.26 % of reading		
	-100 °C to -25 °C	0.27 % of reading		
	-25 °C to 350 °C	0.061 % of reading		
	350 °C to 650 °C	0.03 % of reading		
	650 °C to 1 000 °C	0.017 % of reading		
Electrical Simulation of Temperature Calibration, Indication, and Control Equipment used with Thermocouple Type J ^{FO}	-250 °C to -100 °C	0.14 °C		
	-100 °C to -30 °C	0.27 °C		
	-30 °C to 150 °C	0.24 °C		
	150 °C to 760 °C	0.48 °C		
	760 °C to 1 200 °C	0.19 °C		



Certificate of Accreditation: Supplement

LABCAT, S. de R.L. M.I.

Galeana, No. 619, Col. Moderna

Matamoros, Tamaulipas, México. C.P. 87300

Contact Name: J. Alexandro Santos Vargas Phone: (868) 149-0812

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Electrical Simulation of Temperature Calibration, Indication, and Control Equipment used with Thermocouple Type K ^{FO}	-200 °C to -100 °C	0.17 % of reading	Fluke 5522A Electrical Simulation of Thermocouple Output	ASTM E220
	-100 °C to -25 °C	0.3 % of reading		
	-25 °C to 120 °C	0.14 % of reading		
	120 °C to 1 000 °C	0.07 % of reading		
	1 000 °C to 1 372 °C	0.032 % of reading		
Electrical Simulation of Temperature Calibration, Indication, and Control Equipment used with Thermocouple Type L ^{FO}	-200 °C to -100 °C	0.43 °C		
	100 °C to 800 °C	0.3 °C		
	800 °C to 900 °C	0.21 °C		
Electrical Simulation of Temperature Calibration, Indication, and Control Equipment used with Thermocouple Type N ^{FO}	-200 °C to -100 °C	0.21 % of reading		
	-100 °C to -25 °C	0.36 % of reading		
	-25 °C to 120 °C	0.16 % of reading		
	120 °C to 410 °C	0.05 % of reading		
	410 °C to 1 300 °C	0.022 % of reading		
Electrical Simulation of Temperature Calibration, Indication, and Control Equipment used with Thermocouple Type R ^{FO}	0 °C to 250 °C	0.45 % of reading		
	250 °C to 400 °C	0.11 % of reading		
	400 °C to 1 000 °C	0.52 % of reading		
	1 000 °C to 1 767 °C	0.021 % of reading		
Electrical Simulation of Temperature Calibration, Indication, and Control Equipment used with Thermocouple Type S ^{FO}	0 °C to 250 °C	0.54 °C		
	250 °C to 1 000 °C	0.35 °C		
	1 000 °C to 1 400 °C	0.33 °C		
	1 400 °C to 1 767 °C	0.4 °C		
Electrical Simulation of Temperature Calibration, Indication, and Control Equipment used with Thermocouple Type T ^{FO}	-250 °C to -150 °C	0.33 % of reading		
	-150 °C to 0 °C	0.39 % of reading		
	0 °C to 120 °C	0.14 % of reading		
	120 °C to 400 °C	0.04 % of reading		
Electrical Simulation of Temperature Calibration, Indication, and Control Equipment used with Thermocouple Type U ^{FO}	-200 °C to 0 °C	0.29 °C		
	0 °C to 600 °C	0.54 °C		



Certificate of Accreditation: Supplement

LABCAT, S. de R.L. M.I.

Galeana, No. 619, Col. Moderna

Matamoros, Tamaulipas, México. C.P. 87300

Contact Name: J. Alejandro Santos Vargas Phone: (868) 149-0812

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Equipment to Measure DC Voltage ^{FO}	Up to 329.999 9 mV	0.002 2 % of reading	Fluke 5522A	Euramet-cg-15
	329.999 9 mV to 3.299 999 V	0.000 92 % of reading		
	3.299 999 V to 32.999 99 V	0.001 % of reading		
	32.999 99 V to 329.999 9 V	0.001 4 % of reading		
	329.999 9 V to 1 020 V	0.001 6 % of reading		
Equipment to Measure DC Current ^{FO}	Up to 329.999 μ A	0.016 % of reading	Fluke 5522A/ Coil 50 Turns	
	329.999 μ A to 3.299 99 mA	0.55 % of reading		
	3.299 99 μ A to 32.999 9 mA	0.097 % of reading		
	32.999 9 mA to 329.999 mA	0.011 % of reading		
	329.999 mA to 1.099 99 A	0.02 % of reading		
	1.099 99 A to 2.999 99 A	0.031 % of reading		
	2.999 99 A to 10.999 9 A	0.042 % of reading		
	10.999 9 A to 20.5 A	0.08 % of reading		
	20.5 A to 1 000 A	0.064 % of reading		
Equipment to Measure Resistance ^{FO}	Up to 10.999 9 Ω	0.011 % of reading	Fluke 5522A	
	11 Ω to 32.999 9 Ω	0.038 % of reading		
	33 Ω to 109.999 9 Ω	0.013 % of reading		
	110 Ω to 329.999 9 Ω	0.006 9 % of reading		
	330 k Ω to 1.099 999 k Ω	0.003 6 % of reading		
	1.1 k Ω to 3.299 999 k Ω	0.006 9 % of reading		
	3.3 k Ω to 10.999 99 k Ω	0.002 9 % of reading		
	11 k Ω to 32.999 99 k Ω	0.004 5 % of reading		
	33 k Ω to 109.999 9 k Ω	0.002 9 % of reading		
	110 k Ω to 329.999 99 k Ω	0.004 8 % of reading		
	330 k Ω to 1.099 999 M Ω	0.003 2 % of reading		
	1.1 M Ω to 3.299 999 M Ω	0.008 2 % of reading		
	3.3 M Ω to 10.999 99 M Ω	0.012 % of reading		
	11 M Ω to 32.999 99 M Ω	0.025 % of reading		



Certificate of Accreditation: Supplement

LABCAT, S. de R.L. M.I.

Galeana, No. 619, Col. Moderna

Matamoros, Tamaulipas, México. C.P. 87300

Contact Name: J. Alexandro Santos Vargas Phone: (868) 149-0812

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Equipment to Measure Resistance ^{FO}	33 M Ω to 109.999 9 M Ω	0.041 % of reading	Fluke 5522A	Euramet-cg-15
	110 M Ω to 329.999 9 M Ω	0.26 % of reading		
	330 M Ω to 1 100 M Ω	1.2 % of reading		
Equipment to Measure AC Voltage at listed frequencies ^{FO}				
10 Hz to 45 Hz	1 mV to 32.999 mV	0.18 % of reading		
45 Hz to 10 kHz	1 mV to 32.999 mV	0.17 % of reading		
10 kHz to 20 kHz	1 mV to 32.999 mV	0.038 % of reading		
20 kHz to 50 kHz	1 mV to 32.999 mV	0.094 % of reading		
50 kHz to 100 kHz	1 mV to 32.999 mV	0.3 % of reading		
100 kHz to 500 kHz	1 mV to 32.999 mV	0.75 % of reading		
Equipment to Measure AC Voltage at listed frequencies ^{FO}				
10 Hz to 45 Hz	33 mV to 329.999 mV	0.028 % of reading		
45 Hz to 10 kHz	33 mV to 329.999 mV	0.018 % of reading		
10 kHz to 20 kHz	33 mV to 329.999 mV	0.019 % of reading		
20 kHz to 50 kHz	33 mV to 329.999 mV	0.03 % of reading		
50 kHz to 100 kHz	33 mV to 329.999 mV	0.071 % of reading		
100 kHz to 500 kHz	33 mV to 329.999 mV	0.18 % of reading		
Equipment to Measure AC Voltage at listed frequencies ^{FO}				
10 Hz to 45 Hz	330 mV to 3.299 99 V	0.025 % of reading		
45 Hz to 10 kHz	330 mV to 3.299 99 V	0.013 % of reading		
10 kHz to 20 kHz	330 mV to 3.299 99 V	0.016 % of reading		
20 kHz to 50 kHz	330 mV to 3.299 99 V	0.025 % of reading		
50 kHz to 100 kHz	330 mV to 3.299 99 V	0.059 % of reading		
100 kHz to 500 kHz	330 mV to 3.299 99 V	0.21 % of reading		
Equipment to Measure AC Voltage at listed frequencies ^{FO}				
10 Hz to 45 Hz ^F	3.3 V to 32.9 999 V	0.025 % of reading		
45 Hz to 10 kHz	3.3 V to 32.9 999 V	0.013 % of reading		



Certificate of Accreditation: Supplement

LABCAT, S. de R.L. M.I.

Galeana, No. 619, Col. Moderna

Matamoros, Tamaulipas, México. C.P. 87300

Contact Name: J. Alexandro Santos Vargas Phone: (868) 149-0812

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Equipment to Measure AC Voltage at listed frequencies ^{FO}			Fluke 5522A	Euramet-cg-15
10 kHz to 20 kHz	3.3 V to 32.9 999 V	0.02 % of reading		
20 kHz to 50 kHz	3.3 V to 32.9 999 V	0.03 % of reading		
50 kHz to 100 kHz	3.3 V to 32.9 999 V	0.075 % of reading		
Equipment to measure AC Voltage at listed frequencies ^{FO}				
45 Hz to 1 kHz	33 V to 329.999 V	0.015 % of reading		
1 kHz to 10 kHz	33 V to 329.999 V	0.017 % of reading		
10 kHz to 20 kHz	33 V to 329.999 V	0.021 % of reading		
20 kHz to 50 kHz	33 V to 329.999 V	0.025 % of reading		
50 kHz to 100 kHz	33 V to 329.999 V	0.17 % of reading		
Equipment to measure AC Voltage at listed frequencies ^{FO}				
45 Hz to 1 kHz	330 V to 1 020 V	0.025 % of reading		
1 kHz to 5 kHz	330 V to 1 020 V	0.021 % of reading		
5 kHz to 10 kHz	330 V to 1 020 V	0.025 % of reading		
Equipment to Measure AC Current At listed frequencies ^{FO}				
10 Hz to 20 Hz	29 μ A to 329.99 μ A	0.18 % of reading		
20 Hz to 45 Hz	29 μ A to 329.99 μ A	0.14 % of reading		
45 Hz to 1 kHz	29 μ A to 329.99 μ A	0.12 % of reading		
1 to 5 kHz	29 μ A to 329.99 μ A	0.27 % of reading		
5 kHz to 10 kHz	29 μ A to 329.99 μ A	0.67 % of reading		
10 kHz to 30 kHz	29 μ A to 329.99 μ A	1.4 % of reading		
Equipment to Measure AC Current At listed frequencies ^{FO}				
10 Hz to 20 Hz	330 μ A to 3.299 99 mA	0.16 % of reading		
20 Hz to 45 Hz	330 μ A to 3.299 99 mA	0.1 % of reading		
45 Hz to 1 kHz	330 μ A to 3.299 99 mA	0.084 % of reading		
1 kHz to 5 kHz	330 μ A to 3.299 99 mA	0.16 % of reading		



Certificate of Accreditation: Supplement

LABCAT, S. de R.L. M.I.

Galeana, No. 619, Col. Moderna

Matamoros, Tamaulipas, México. C.P. 87300

Contact Name: J. Alexandro Santos Vargas Phone: (868) 149-0812

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Equipment to Measure AC Current At listed frequencies ^{FO}			Fluke 5522A	Euramet-cg-15
5 kHz to 10 kHz	330 μ A to 3.299 99 mA	0.4 % of reading		
10 kHz to 30 kHz	330 μ A to 3.299 99 mA	0.79 % of reading		
Equipment to Measure AC Current At listed frequencies ^{FO}				
10 Hz to 20 Hz	3.3 mA to 32.9 999 mA	0.19 % of reading		
20 Hz to 45 Hz	3.3 mA to 32.9 999 mA	0.14 % of reading		
45 Hz to 1 kHz	3.3 mA to 32.9 999 mA	0.13 % of reading		
1 kHz to 5 kHz	3.3 mA to 32.9 999 mA	0.14 % of reading		
5 kHz to 10 kHz	3.3 mA to 32.9 999 mA	0.2 % of reading		
10 kHz to 30 kHz	3.3 mA to 32.9 999 mA	0.34 % of reading		
Equipment to Measure AC Current At listed frequencies ^{FO}				
10 Hz to 20 Hz	33 mA to 329.999 mA	0.073 % of reading		
20 Hz to 45 Hz	33 mA to 329.999 mA	0.076 % of reading		
45 Hz to 1 kHz	33 mA to 329.999 mA	0.039 % of reading		
1 kHz to 5 kHz	33 mA to 329.999 mA	0.091 % of reading		
5 kHz to 10 kHz	33 mA to 329.999 mA	0.18 % of reading		
10 kHz to 30 kHz	33 mA to 329.999 mA	0.36 % of reading		
Equipment to Measure AC Current At listed frequencies ^{FO}				
10 kHz to 45 Hz	330 mA to 1.099 99 A	0.15 % of reading		
45 Hz to 1 kHz	330 mA to 1.099 99 A	0.047 % of reading		
1 kHz to 5 kHz	330 mA to 1.099 99 A	0.54 % of reading		
5 kHz to 10 kHz	330 mA to 1.099 99 A	2.3 % of reading		
Equipment to Measure AC Current At listed frequencies ^{FO}				
45 Hz to 100 Hz	3 A to 10.999 9 A	0.062 % of reading		
100 Hz to 1 kHz	3 A to 10.999 9 A	0.093 % of reading		
1 kHz to 5 kHz	3 A to 10.999 9 A	2.4 % of reading		



Certificate of Accreditation: Supplement

LABCAT, S. de R.L. M.I.

Galeana, No. 619, Col. Moderna

Matamoros, Tamaulipas, México. C.P. 87300

Contact Name: J. Alexandro Santos Vargas Phone: (868) 149-0812

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Equipment to Measure AC Current At listed frequencies ^{FO}			Fluke 5522A	Euramet-cg-15
45 Hz to 100 Hz	11 A to 20.5 A	0.11 % of reading		
100 Hz to 1 kHz	11 A to 20.5 A	0.14 % of reading		
1 kHz to 5 kHz	11 A to 20.5 A	2.4 % of reading		
Equipment to Measure AC Current At listed frequencies ^{FO}			Fluke 5522A Coil 50 Turns	
45 Hz to 65 Hz	20.5 A to 1 000 A	0.064 % of reading		
Equipment to Measure Capacitance At listed frequencies ^{FO}			Fluke 5522A	
10 Hz to 10 kHz	220 pF to 399.9 pF	2.3 % of reading		
10 Hz to 10 kHz	400 nF to 1.099 9 nF	1.1 % of reading		
10 Hz to 3 kHz	1.1 nF to 3.299 9 nF	0.61 % of reading		
10 Hz to 1 kHz	3.3 nF to 10.999 9 nF	0.26 % of reading		
10 Hz to 1 kHz	11 nF to 32.999 9 nF	0.22 % of reading		
10 Hz to 1 kHz	33 nF to 109.999 nF	0.2 % of reading		
10 Hz to 1 kHz	110 nF to 329.999 nF	0.21 % of reading		
10 Hz to 600 Hz	330 μ F to 1.099 99 μ F	0.2 % of reading		
10 Hz to 300 Hz	1.1 μ F to 3.299 99 μ F	0.26 % of reading		
10 Hz to 150 Hz	3.3 μ F to 10.999 9 μ F	0.28 % of reading		
10 Hz to 120 Hz	11 μ F to 32.999 9 μ F	0.4 % of reading		
10 Hz to 80 Hz	33 μ F to 109.999 μ F	0.43 % of reading		
Equipment to Measure Capacitance At listed frequencies ^{FO}				
0 Hz to 50 Hz	110 μ F to 329.999 μ F	0.44 % of reading		
0 Hz to 20 Hz	330 μ F to 1.099 99 mF	0.42 % of reading		
0 Hz to 6 Hz	1.1 mF to 3.299 99 mF	0.42 % of reading		
0 Hz to 2 Hz	3.3 mF to 10.999 9 mF	0.42 % of reading		
0 Hz to 0.6 Hz	11 mF to 32.999 9 mF	0.65 % of reading		
0 Hz to 0.2 Hz	33 mF to 110 mF	0.92 % of reading		



Certificate of Accreditation: Supplement

LABCAT, S. de R.L. M.I.

Galeana, No. 619, Col. Moderna

Matamoros, Tamaulipas, México. C.P. 87300

Contact Name: J. Alexandro Santos Vargas Phone: (868) 149-0812

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED		
Equipment to Output DC Voltage ^{FO}	100 mV	0.003 7 % of reading + 0.003 5 % range	6 ½ Multimeter Fluke 8846A	Euramet cg-15		
	1 V	0.002 5 % of reading + 0.000 7 % range				
	10 V	0.002 4 % of reading + 0.000 5 % range				
	100 V	0.003 8 % of reading + 0.000 6 % range				
	1 000 V	0.004 1 % of reading + 0.001 % range				
Equipment to Output AC Voltage At the listed frequencies ^{FO}						
3 Hz to 5 Hz	100 mV	1 % of reading + 0.04% range				
5 Hz to 10 Hz	100 mV	0.35 % of reading + 0.04% range				
10 Hz to 20 kHz	100 mV	0.06 % of reading + 0.04 % range				
20 kHz to 50 kHz	100 mV	0.12 % of reading + 0.05 % range				
50 kHz to 100 kHz	100 mV	0.6 % of reading g + 0.08 % range				
100 kHz to 300 kHz	100 mV	4 % of reading + 0.5 % range				
Equipment to Output AC Voltage At the listed frequencies ^{FO}						
3 Hz to 5 Hz	1 V	1 % of reading + 0.03 % range				
5 Hz to 10 Hz	1 V	0.35 % of reading + 0.03 % range				
10 Hz to 20 kHz	1 V	0.06 % of reading + 0.03 % range				
20 kHz to 50 kHz	1 V	0.12 % of reading + 0.05 % range				
50 kHz to 100 kHz	1 V	0.6 % of reading + 0.08 % range				
100 kHz to 300 kHz	1 V	4 % of reading + 0.5 % range				
Equipment to Output AC Voltage At the listed frequencies ^{FO}						
3 Hz to 5 Hz	10 V	1 % of reading + 0.03 % range				
5 Hz to 10 Hz	10 V	0.35 % of reading + 0.03 % range				
10 Hz to 20 kHz	10 V	0.06 % of reading dg + 0.03 % range				
20 kHz to 50 kHz	10 V	0.12 % of reading + 0.05 % range				
50 kHz to 100 kHz	10 V	0.6 % of reading + 0.08 % range				
100 kHz to 300 kHz	10 V	4 % of reading + 0.5 % range				



Certificate of Accreditation: Supplement

LABCAT, S. de R.L. M.I.

Galeana, No. 619, Col. Moderna

Matamoros, Tamaulipas, México. C.P. 87300

Contact Name: J. Alexandro Santos Vargas Phone: (868) 149-0812

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Equipment to Output AC Voltage At the listed frequencies ^{FO}			6 ½ Multimeter Fluke 8846A	Euramet cg-15
3 Hz to 5 Hz	100 V	1 % of reading dg + 0.03 % range		
5 Hz to 10 Hz	100 V	0.35 % of reading + 0.03 % range		
10 Hz to 20 kHz	100 V	0.06 % of reading + 0.03 % range		
20 kHz to 50 kHz	100 V	0.12 % of reading + 0.05 % range		
50 kHz to 100 kHz	100 V	0.6 % of reading + 0.08 % range		
Equipment to Output AC Voltage At the listed frequencies ^{FO}				
3 Hz to 5 Hz	1 000 V	1 % of reading + 0.023 % range		
5 Hz to 10 Hz	1 000 V	0.35 % of reading + 0.023 % range		
10 Hz to 10 kHz	1 000 V	0.06 % of reading + 0.023 % range		
Equipment to Output DC Current ^{FO}			6 ½ Multimeter Fluke 8846A Current Shunt	Euramet cg-15
100 µA		0.05 % of reading + 0.025 % range		
1 mA		0.05 % of reading g + 0.005 % range		
10 mA		0.05 % of reading + 0.02 % range		
100 mA		0.05 % of reading + 0.005 % range		
400 mA		0.05 % of reading + 0.005 % range		
1 A		0.05 % of reading + 0.02 % range		
3 A		1 % of reading + 0.02 % range		
10 A		0.15 % of reading + 0.008 % range		
Equipment to Output DC Current (Indirect Method)	100 A	0.004 9 % of reading		
Equipment to Output AC Current At the listed frequencies ^{FO}			6 ½ Multimeter Fluke 8846A	
5 Hz to 10 Hz	100 µA	1.1 % of reading + 0.06 % range		
10 Hz to 5 kHz	100 µA	0.35 % of reading + 0.06 % range		
5 kHz to 10 kHz	100 µA	0.15 % of reading + 0.06 % range		
Equipment to Output AC Current At the listed frequencies ^{FO}				
5 Hz to 10 Hz	1 mA	0.3 % of reading + 0.04 % range		
10 Hz to 5 kHz	1 mA	0.1 % of reading + 0.04 % range		



Certificate of Accreditation: Supplement

LABCAT, S. de R.L. M.I.

Galeana, No. 619, Col. Moderna

Matamoros, Tamaulipas, México. C.P. 87300

Contact Name: J. Alejandro Santos Vargas Phone: (868) 149-0812

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Equipment to Output AC Current At the listed frequencies ^{FO}			6 ½ Multimeter Fluke 8846A	Euramet cg-15
5 kHz to 10 kHz	1 mA	0.2 % of reading + 0.25 % range		
Equipment to Output AC Current At the listed frequencies ^{FO}				
5 Hz to 10 Hz	10 mA	0.35 % of reading + 0.06 % range		
10 Hz to 5 kHz	10 mA	0.15 % of reading + 0.06 % range		
5 kHz to 10 kHz	10 mA	0.35 % of reading + 0.7 % range		
Equipment to Output AC Current At the listed frequencies ^{FO}				
5 Hz to 10 Hz	100 mA	0.3 % of reading dg + 0.04 % range		
10 Hz to 5 kHz	100 mA	0.1 % of reading + 0.04 % range		
5 kHz to 10 kHz	100 mA	0.2 % of reading + 0.25 % range		
Equipment to Output AC Current At the listed frequencies ^{FO}				
5 Hz to 10 Hz	400 mA	0.3 % of reading + 0.1 % range		
10 Hz to 5 kHz	400 mA	0.1 % of reading + 0.1 % range		
5 kHz to 10 kHz	400 mA	0.2 % of reading + 0.7 % range		
Equipment to Output AC Current At the listed frequencies ^{FO}				
10 Hz to 45 Hz	1 A	0.1 % of reading + 0.04 % range		
45 Hz to 5 kHz	1 A	0.1 % of reading + 0.04 % range		
5 kHz to 10 kHz	1 A	0.35 % of reading + 0.7 % range		
Equipment to Output AC Current At the listed frequencies ^{FO}				
45 Hz to 100 Hz	3 A	0.15 % of reading + 0.06 % range		
100 Hz to 1 kHz	3 A	0.15 % of reading + 0.06 % range		
1 kHz to 10 kHz	3 A	0.35 % of reading + 0.7 % range		



Certificate of Accreditation: Supplement

LABCAT, S. de R.L. M.I.

Galeana, No. 619, Col. Moderna

Matamoros, Tamaulipas, México. C.P. 87300

Contact Name: J. Alexandro Santos Vargas Phone: (868) 149-0812

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Equipment to Output AC Current At the listed frequencies ^{FO}			6 ½ Multimeter Fluke 8846A	Euramet cg-15
45 Hz to 100 Hz	10 A	0.15 % of reading + 0.06 % range		
100 Hz to 1 kHz	10 A	0.15 % of reading + 0.06 % range		
1 kHz to 10 kHz	10 A	0.35 % of reading + 0.7 % range		
Equipment to Output AC Current At the listed frequencies (Indirect Method) ^{FO}			6 ½ Multimeter Fluke 8846A Current Shunt	Euramet cg-15
45 Hz to 100 Hz	100 A	0.13 % of reading		
100 Hz to 1 kHz	100 A	0.16 % of reading		
Equipment to Output Electrical Resistance ^{FO}	10 Ω	0.01 % of reading + 0.03 % range	6 ½ Multimeter Fluke 8846A	
	100 Ω	0.01 % of reading + 0.004 % range		
	1 k Ω	0.01 % of reading + 0.001 % range		
	10 k Ω	0.01 % of reading + 0.001 % range		
	100 k Ω	0.01 % of reading + 0.001 % range		
	1 M Ω	0.01 % of reading + 0.001 % range		
	10 M Ω	0.04 % of reading + 0.001 % range		
	100 M Ω	0.01 % of reading + 0.01 % range		
	1 000 M Ω	2 % of reading + 0.01v% range		
Equipment to Output Capacitance ^{FO}	10 nF	2 % of reading + 2.5 % range		
	100 nF	1 % of reading + 0.5 % range		
	1 μ F	1 % of reading + 0.5 % range		
	10 μ F	1 % of reading + 0.5 % range		
	100 μ F	1 % of reading + 0.5 % range		
	1 mF	1 % of reading + 0.5 % range		
	10 mF	1 % of reading + 0.5 % range		
	100 mF	4 % of reading + 0.2 % range		



Certificate of Accreditation: Supplement

LABCAT, S. de R.L. M.I.

Galeana, No. 619, Col. Moderna

Matamoros, Tamaulipas, México. C.P. 87300

Contact Name: J. Alexandro Santos Vargas Phone: (868) 149-0812

Accreditation is granted to the facility to perform the following calibrations:

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its fixed location. Example: Outside Micrometer^F would mean that the laboratory performs this calibration at its fixed location.
4. The presence of a superscript O means that the laboratory performs calibration of the indicated parameter onsite at customer locations. Example: Outside Micrometer^O would mean that the laboratory performs this calibration onsite at the customer's location.
5. The presence of a superscript FO means that the laboratory performs calibration of the indicated parameter both at its fixed location and onsite at customer locations. Example: Outside Micrometer^{FO} would mean that the laboratory performs this calibration at its fixed location and onsite at customer locations.
6. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.
7. The term L represents length in inches or millimeters as appropriate to the uncertainty statement.
8. The term T represents temperature in °C or °F as appropriate to the uncertainty statement.
9. The term Wt represents weight in pounds or grams (including SI multiple and submultiple units) appropriate to the uncertainty statement.